



CERTIFICATE OF ANALYSIS

Product Name	SDA w/ Gentamycin + Chloramphenicol
Catalogue No	02051
Batch No	2303016
Dehydrated Brand / Batch No / Expiry Date	SDA: Oxoid/3431660/02-2027
Neutralizer Brand / Batch No / Expiry Date	Gentamycin (İbrahim Etem): 2012003 / 12.2023 Chloramphenicol (Glentham): 545BHK / 12.2023
Production Date	06.03.2023
Expiry Date	06.06.2023
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Must have a smooth surface	Appropriate
Color	Amber	Appropriate
pH	5,6 ± 0,2	5,63

Microbiological Specifications

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 25±2 °C:48h-5d	
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)			Sterile	
QC Strains	ATCC	Results			
		Recovery		Growth	
Candida albicans	10231	75	73	Good	Appropriate
Aspergillus brasiliensis	16404	73	72	Good	Appropriate
Escherichia coli	25922	Inhibition	Inhibition	Inhibition	Appropriate
S.aureus	25923	Inhibition	Inhibition	Inhibition	Appropriate

Acceptance Criteria: %50-200

QC Date	13.03.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net