



CERTIFICATE OF ANALYSIS

Product Name	Reinforced Clostridial Medium (100 ml)
Catalogue No	01124
Batch No	2303022
Dehydrated Brand / Batch No / Expiry Date	Reinforced Clostridial Medium : Condalab/ 907163/ 07-2023
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	06.03.2023
Expiry Date	06.09.2023
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Opalescent with particles	Appropriate
Color	Light to medium amber	Appropriate
pH	6,8 ± 0,2	6,85

Microbiological Specifications

Method of control: Direct Inoculation	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)		Incubation: 35±2 °C:24-48 h 25±2 °C:6 d		
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)			Sterile	
QC Strains	ATCC	Results			
		Recovery		Growth	
Clostridium perfringens	13124	Good	Good	Good	Appropriate
Escherichia coli	25922	Good	Good	Good	Appropriate
Staphylococcus aureus	25923	Good	Good	Good	Appropriate
Pseudomonas aeruginosa	27853	Inhibition	Inhibition	Inhibition	Appropriate

QC Date	13.03.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net