



CERTIFICATE OF ANALYSIS

Product Name	TSB STERILITY TEST BROTH (100mL)				
Catalogue No	01207				
Batch No	2303048				
Dehydrated Brand / Batch No / Expiry Date	Tryptic Soy Broth :/ 0000526203/ 02-2027				
Neutralizer Brand / Batch No / Expiry Date	-				
Production Date	14.03.2023				
Expiry Date	14.09..2023				
Storage Conditions	2-25°C				
Physical & Chemical Specifications					
Appearance	Clear			Appropriate	
Color	Yellow			Appropriate	
pH	7,3 ± 0,2			7,34	
Microbiological Specifications					
Method of control: Direct Inoculation	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 35±2 °C:24-48 h 25±2 °C:6 d	
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)			Sterile	
QC Strains	ATCC	Results			
		Recovery		Growth	
Bacillus subtilis	6633	Good	Good	Good	Appropriate
Staphylococcus aureus	6538	Good	Good	Good	Appropriate
Candida albicans	10231	Good	Good	Good	Appropriate
Pseudomonas aeruginosa	9027	Good	Good	Good	Appropriate
Aspergillus brasiliensis	16404	Good	Good	Good	Appropriate
QC Date	21.03.2023				
QC Manager	Ali ARAS				

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net