



CERTIFICATE OF ANALYSIS

Product Name	%5 Sheep Blood Agar
Catalogue No	02001
Batch No	2303060
Dehydrated Brand / Batch No / Expiry Date	CBA: Oxoid/3554393/10-2027
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	20.03.2023
Expiry Date	20.06.2023
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Must have a smooth surface	Appropriate
Color	Red	Appropriate
pH	7,4 ± 0,2	7,43

Microbiological Specifications

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 30-35 °C: 18-24h	
Sterility	Incubation: (30-35°C : 4d)			Sterile	
QC Strains	ATCC	Results			
		Recovery		Growth	
Escherichia coli	25922	72	74	Good (γ hemolysis)	Appropriate
Staphylococcus aureus	25923	73	74	Good (β hemolysis)	Appropriate
Staphylococcus pyogenes	19615	73	73	Good (β hemolysis)	Appropriate
Staphylococcus pneumoniae	6305	72	73	Good (α hemolysis)	Appropriate

Acceptance Criteria: %50-200

QC Date	27.03.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net