



CERTIFICATE OF ANALYSIS

Product Name	Mueller Hinton Agar
Catalogue No	02033
Batch No	2303064
Dehydrated Brand / Batch No / Expiry Date	MHA: Condalab/303011/ 03-2026
Neutralizer Brand / Batch No / Expiry Date	-
Production Date	21.03.2023
Expiry Date	21.09.2023
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	must have a smooth surface	Appropriate
Color	Light amber	Appropriate
pH	7,3 ± 0,2	7,35

Microbiological Specifications

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)		Incubation: 35±2 °C:24-48h		
Sterility	Incubation: (20-25°C : 3 d, 30-35°C : 2 d)		Sterile		
QC Strains	ATCC	Results			
		Control	Product	Growth	
Escherichia coli	25922	75	74	Good	Appropriate
Staphylococcus aureus	25923	74	73	Good	Appropriate
Pseudomanas aeruginosa	27853	74	75	Good	Appropriate
Enterococcus faecalis	29212	73	74	Good	Appropriate

Acceptance Criteria: %50-200

QC Date	28.03.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net