



CERTIFICATE OF ANALYSIS

Product Name	Enterococcosel Agar w / Vancomycin
Catalogue No	02026
Batch No	2311086
Dehydrated Brand / Batch No / Expiry Date	Enterococcosel Agar: Condalab/ 309204/09-2026
Neutralizer Brand / Batch No / Expiry Date	Vancomycin HCL / 0000130509 / 02-2028
Production Date	20.11.2023
Expiry Date	20.02.2024
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Must have a smooth surface	Appropriate
Color	Dark green	Appropriate
pH	7,1 ± 0,2	7,12

Microbiological Specifications

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)		Incubation: 30-35 °C: 24-48h		
Sterility	Incubation: (20-25°C : 3d, 30-35°C : 2d)			Sterile	
QC Strains	ATCC	Results			
		Control	Product	Growth	
Enterococcus faecalis	51299	73	74	Brown-Black Colonies	Appropriate
Enterococcus faecalis	29212	Partial Inhibition	Partial Inhibition	Partial Inhibition	Appropriate
Escherichia coli	25922	Partial Inhibition	Partial Inhibition	Partial Inhibition	Appropriate
Streptococcus pyogenes	19615	Partial Inhibition	Partial Inhibition	Partial Inhibition	Appropriate

Acceptance Criteria: %50-200

QC Date	27.11.2023
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net