



CERTIFICATE OF ANALYSIS

Product Name	Endo Agar
Catalogue No	02024
Batch No	2409037
Dehydrated Brand / Batch No / Expiry Date	Endo Agar: Condalab/ 405192 / 05-2027
Neutralizer Brand / Batch No / Expiry Date	Basic Fuchsin(Himedia): 0000470458 / 02-2025 Ethanol(Scharlau): 21449610 / 01-2026
Production Date	13.09.2024
Expiry Date	13.12.2024
Storage Conditions	2-12°C

Physical & Chemical Specifications

Appearance	Must have a smooth surface	Appropriate
Color	Clear and pale pink	Appropriate
pH	7,4 ± 0,2	7,43

Microbiological Specifications

Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)	Incubation: 35±2 °C:18-48h
Sterility	Incubation: (20-25°C : 3 d, 30-35°C : 2 d)	Sterile
QC Strains	ATCC	Results
		ControlProductGrowth
Escherichia coli	8739	7373ReddishAppropriate
Salmonella typhimurium	14028	7575ColourlessAppropriate
Enterococcus faecalis	29212	InhibitionInhibitionInhibitionAppropriate

Acceptance Criteria: %50-200

QC Date	16.09.2024
QC Manager	Ali ARAS

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net