



CERTIFICATE OF ANALYSIS

Product Name	Eosin Methylene Blue Agar				
Catalogue No	02023				
Batch No	2501031				
Dehydrated Brand / Batch No / Expiry Date	Eosin Methylene Blue Agar : Oxoid / 3718572/08-2028				
Neutralizer Brand / Batch No / Expiry Date	-				
Production Date	09.01.2025				
Expiry Date	09.07.2025				
Storage Conditions	2-12°C				
Physical & Chemical Specifications					
Appearance	Must have a smooth surface			Appropriate	
Color	Reddish purple			Appropriate	
pH	6,8±0,2			6,83	
Microbiological Specifications: Productivity and acc. To ISO/TS 11133					
Method of control: Counting Method	Inoculum: (+ control 10-100 cfu, - control 100 - 1000 cfu)			Incubation: 30-35 °C: 24-48h	
Sterility	Incubation: (20-25°C : 3 d, 30-35°C : 2 d)			Sterile	
QC Strains	ATCC	Results			
		Control	Product	Growth	
Escherichia coli	25922	73	75	Metallic shine	Appropriate
Salmonella typhimurium	14028	74	74	Colourless	Appropriate
Enterococcus faecalis	29212	Partial Inhibition	Partial Inhibition	Inhibition	Appropriate
Staphylococcus aureus	25923	Inhibition	Inhibition	Inhibition	Appropriate
Acceptance Criteria:%50-200					
QC Date	13.01.2025				
QC Manager	Ali ARAS				

This document has been prepared electronically.

BİOCELL BİYOLOJİK VE TIBBİ ÜRÜNLER SAN. TİC. LTD. ŞTİ.
Kaynarca Mahallesi Çeşni sokak No : 5/2 Pendik / İstanbul / TURKEY
Telefon:0216 354 90 54 E-mail: info@bio-cell.net